

Peter Vock

## CT dose reduction in children

---

Published online: 10 September 2005  
© Springer-Verlag 2005

---

**Eur Radiol (2005) s00330-005-2856-0**

Unfortunately, there were errors in Tables 3 and 5. The corrected tables are given here.

---

The online version of the original article can be found at  
<http://dx.doi.org/10.1007/s00330-005-2856-0>

---

P. Vock (✉)  
Department of Radiology,  
University Hospital Inselspital,  
3010 Bern, Switzerland  
e-mail: [peter.vock@insel.ch](mailto:peter.vock@insel.ch)  
Tel.: +41-31-6322435  
Fax: +41-31-6324874

**Table 3** Suggested paediatric CT protocols. *CTDI* CT dose index, *DLP* dose-length product, *ED* effective dose

Weight (kg)	CTDI	kV	mAs	Rows	Comment	Reference
4.5 –<9/9 –<18			40/50	4	Chest and abdomen	[10, 13]
18 –<27/27 –<36			60/70			
36–45			80			
>45–69			100–120			
>70			≥140			
2.5–5	7.1	80	72	4×2.5	Abdomen pitch 0.75	[42]
5–15	9.4	100	56			
15–30	14.0	120	64			
30–50	18.5	120	96			
<15			15%	4	Chest/abdomen % of adult values	[40]
15–24			25%			
25–34			40/35%			
35–44			60/50%			
45–54			80/75%			
>54			100%			
Age (years)	CTDI		DLP	ED (mSV)	Comment	Reference
<1	25/20		180/150	1/5	Brain/chest	[30]
5	25/25		200/200	1/5		
10	50/30		750/600	2/5		
<1	20		330/170	4–8	Upper/lower	
5	25		360/250	4–8	abdomen	
10	30		800/500	5–12		

**Table 5** Effective dose estimated from dose-length product (DLP)<sup>a</sup> [8, 38]

Age	Head	Neck	Chest	Abdomen/pelvis
0 y	0.011/0.027	0.017/-	0.039/0.034	0.049/0.040
1 y	0.007/0.008	0.012/-	0.026/0.021	0.030/0.024
5 y	0.004/0.004	0.011/-	0.018/0.014	0.020/0.016
10 y	0.003/0.003	0.008/-	0.013/0.011	0.0150/0.014
15 y			-/0.015	-/0.009
Adult	0.002/0.003	0.006/-	0.014/0.009	0.015/-

<sup>a</sup>Numbers give normalised effective dose per dose-length product (mSv per mGy.cm)